

DRY BATH INCUBATOR LTBAT23-26



DRY BATH INCUBATOR LTBAT23-26

Used in Clinical, General Chemistry, Preservation, Reactions of sample, DNA amplification, Initial Denaturation of electrophoresis.

Also known as Laboratory Dry Bath Chillers.

LTBAT23-26 DRY BATH INCUBATOR

Fast heating speed, uniform heating, accurate temperature control, high stability, low energy consumption and no noise.

Simple or program temperature control mode is available, making the experiment more convenient.

Sliding operation, light touch button, novel and fashionable.

The standard metal blocks are easy to replace, easy to clean and disinfect.

Protect samples from contamination.

The buzzer tone can be turned off, making the experiment quieter.



SPECIFICATIONS

Model	LTBAT23-26
Temperature Range	RT+5-150°C
Temp. Setting Range	5°C~150°C
Time Range	1min~99h59m (00:00 is continuous)
Block Temp. Stability@40~100°C	±0.5 °C
Block Temp. Stability@>100°C	±1 °C
Block Temp. Uniformity@40°C	±0.3 °C
Block Temp. Uniformity@>40°C	±0.5 °C
Temp. Display Accuracy	0.1 °C
Heating Speed	≤15min(20°C~150°C)
Sample Capacity	2 standard block
Voltage	AC 220V/AC 110V 50Hz/60Hz
Power	400W
Fuse	250V,3A/6A,φ5x20
Dimension	W.285xD.225xH.95
Net Weight(kgs)	2.64 kgs

LABTARE

Labtare Analytical Instruments

Email: info@labtare.com | Website: labtare.com