

GAS CHROMATOGRAPHY SYSTEM LTCGR8-7



GAS CHROMATOGRAPHY SYSTEM LTCGR8-7

Used in Food Testing, Chemical Industry, Beverage Testing, Drug testing, Forensic Science, Pharmaceutical, Molecular Biology, Medical, Research, Laboratory.

LTCGR8-7 GAS CHROMATOGRAPHY SYSTEM

PC control, user-friendly interface, and easy to operate.

Heating speed is fast and overshoot temperature is small.

Self-diagnosis, power protection, oven over-temperature protection, and automatic ignition.

It can accurately display the temperature control settings, actual value, and FID amplifier sensitivity.

The single gas system and precise scale pneumatic control valve contribute to excellent reproducibility and stability and can perform analysis of packed column or capillary with wide diameter of 0.53mm.

Packed columns: On-column injection, instantaneous vaporization injection, gas injection.

Open computer system and chromatography workstation can work together to process data.

Large capacity oven facilitates the installation of packed column and capillary.

Built-in heating wire structure.

RS232 communication port.



SPECIFICATIONS

Model	LTCGR8-7
Column Oven	
Temperature Range	15°C – 399°C above room temperature (increment: 1°C)
Temperature Control Accuracy	Better than $\pm 0.1^\circ\text{C}$ (measured at 200°C)
Hydrogen flame ionization detector (FID)	
Detection limit	$\text{Dt} \leq 1 \times 10^{-10} \text{ g/s}$ (octane and hexadecane)
Baseline drift	$\leq 2 \times 10^{-12} \text{ A/h}$
Linear range	$\geq 10^6$
Max. limit temperature	400°C
Other Specifications	
Dimension (LxWxH)	575x480x490 mm
Weight (Kg)	50
Power supply voltage	220 V - ± 22 V 50 Hz ± 0.5 Hz
Power	$\leq 1500 \text{ W}$



Labtare Analytical Instruments

Email: info@labtare.com | Website: labtare.com