

# SPIN TISSUE PROCESSOR LTHIS10-5



# SPIN TISSUE PROCESSOR LTHIS10-5

Used in Clinical and Research Histopathology.  
Also known as Tissue Embedding Cassettes.

## LTHIS10-5 SPIN TISSUE PROCESSOR

Adopt Japan Mitsubishi PLC to control the whole working process, easy to operate, stable and

reliable work. If lose electric give an alarm,protecting tissue in liquid,obstache,protect function and so on. Adopting agitation dehydrate mode,so tissue and solvent,paraffin meet,thereby well effect. Time setting range It can be turned on any day or every day, the working time of each cylinder

can be up to 999 minutes, the minimum is 1 minute for the 1st and 12th cylinders, and the

minimum is 0 for the rest. Adopting 9 slices 1.2Lmedicine aquarium,at any moment observe tissue change at work. It can be equipped with power supply according to user requirements, and can operate for 4-16

hours for other functions besides wax bath after power failure to ensure the safety of the

Organization



## SPECIFICATIONS

Model	LTHIS10-5
Steps of processing tissue	12 steps
Glass container	Quantity — 9 capacity — 1.2 L
Wax cup	Quantity — 3capacity — 1L.
Temperature range	60-99 °C (can be designed according to user requirements)
Process	Each cup working hours — 0~59 hr 59 min.
Conditions of use Ambient temperature	+5°C ~ 40
Conditions of use Ambient humidity	< 85 % (at 20 ± 5 °C)
Vetically reciprocative motion of tissue basket	3 times / min
Tissue basket specification Barrel tissue basket	∅ 95x100 mm
Layer nacelle	∅ 95x80 mm Five layers
Power	400~500 W
Rated voltage	AC 220 V 50 Hz /AC 110 V 50 Hz

# LABTARE

**Labtare Analytical Instruments**

Email: [info@labtare.com](mailto:info@labtare.com) | Website: [labtare.com](http://labtare.com)