

NUCLEIC ACID EXTRACTION SYSTEM LTNAP9-7



NUCLEIC ACID EXTRACTION SYSTEM LTNAP9-7

Used in DNA and RNA Purification, Cultured Cells, Bacteria, Tissues, Cell-Free Body Fluids, Plant Samples, Blotting, PCR, Cloning, Medical Sciences.

Also known as Nucleic acid Extractor.

LTNAP9-7 NUCLEIC ACID EXTRACTION SYSTEM

7-inch touch screen, easy to use, fast response

User-defined cracking and elution temperature

UV disinfection function, time range 1min-24hour

Automatic control system, no need connect to computer

Free programming to meet the needs of different reagent

Open system, fully automatic, stable results and good repeatability

Extract rapidly 9-40 minutes , 32/48 samples can be extracted at the same time



SPECIFICATIONS

Model	LTNAP9-7
Sample Quantity	48
Processing Volume	60 μ L-1000 μ L
Sample Volume	20-500 μ L
Sample Throughput	48
Magnetic Bead Recovery	>98%
Extracting the Difference Between Holes	CV \leq 3%
Heating Temperature	8 independent heating modules, customize lysis and elution temperature (temperature range) according to your needs
Oscillating Mixing	Low,medium and high third gears are adjustable, and the fluctuation range can be adjusted with the reagent volume
Reagent Type	Magnetic bead open platform
Extraction Time	8-60 min/round (depending on the reagent used)
Internal Program	5000 groups
Program Management	Powerful program editing capabilities to meet different reagent needs. U disk program import and export can be achieved
Safety Door Design	After the safety door is opened,the program operation will be automatically suspended, and the program can continue t run after the safety door closed
Built-in Air Duct	Yes
Ultraviolet Irradiation	Yes
Packing Size	700x520x750 mm
Gross Weight(kg)	80 kg



Labtare Analytical Instruments

Email: info@labtare.com | Website: labtare.com