SYNTHESIZING ULTRAPURE WATER SYSTEM WPS67-060FV

6666

labtare.com

SYNTHESIZING ULTRAPURE WATER SYSTEM WPS67-060FV

Ultrapure water system is sub-economic choice for high grade experiments. This level of purification is required for advanced analytical techniques, such as HPLC, and is commonly used for semi-conductor manufacturing. Used in Laboratory, Manufacturing, Reefkeeping, Aquarium, Laboratory, Research.

Also known as Laboratory Ultrapure water system.

WPS67-060FV SYNTHESIZING ULTRAPURE WATER SYSTEM

With tap water inlet, to produce RO water and ultrapure water, quality can reach to 18.2 M Ω .cm.

Built-in 20 liters airtight plastic pressure water tank

Built-in 13 liters high-capacity polishing resin cartridge

Unique design and easy-to-replace cartridges pack unit.

Data storage and RS 232/USB communication port.

3 way on-line water quality sensor, multiple alarm.

Life-span of cartridges' display and alarm.

System circulation function, system sterilization procedure.

Molding process, high-strength, streamline plastic shell.

The graphic display clearly indicates all system's parameters. From water quality to knowing when it is time to change the purification pack, you'll see at a glance what is need

For ease-of-use, the main purification technologies are contained in an innovative allin-one pack that mean you can change it in just a couple of minutes.

The system requires no special installation, connect the system to your tap water supply it's ready to use.

SPECIFICATIONS

Model	WPS67-060FV
Feed Water Requirements*	
Water Inlet	Tap water
Temperature	5-45°C
Pressure	1.0-4.0 Kgf/cm ²
Bacteria	<0.1 cfu/ml
DimensionLxWxH	570x600x1500 mm
Weight	60 kg
Power Consumption (W)	120 W
Power Supply	AC110-220 V, 50/60 Hz
Note	*The quality of output water accords with the quality of inlet water.
Ultrapure Water Quality	
Heavy Metal Ion	<0.1 ppb
Endotoxin	<0.001 EU/ml
Rnases	<0.01 ng/ml
Dnases	<4pg/μl
Feed Water Requirements	



Output	60 L/hrs at 25°C
Flow rate (with pressure tank)	-
Resistivity (25°C)	18.2 MΩ.cm
TOC*	3 ррb
Particle (>0.1µm)	<1/ml
Conductivity of RO water quality	< tap waterx4%



Labtare USA

82 Wendell Avenue, STE 100, Pittsfield, MA, 01201, USA Email: info@labtare.com | Website: labtare.com